

DECLARATION OF CONFORMITY (DOC)

Manufacturer/EU Representative:

Ferno-Washington, Inc. 70 Weil Way Wilmington, Ohio 45177-9371 U.S.A. 1.937.382.1451

EC REP FERNO S.A. via B. Zallone, n. 26, 40066 Pieve di Cento, Bologna, Italy +39.051.6860028

Trade Name:

FERNO®

SRN: N/A

Item/Catalogue #	Item Description	UDI-DI Number (GTIN)	Risk Class
0313778	678 PEDI-MATE	Not implemented to date	ı
0314115	PEDI MATE PLUS	Not implemented to date	T

Intended Use of Medical Device:

A restraint to be attached to a backboard or stretcher in order to immobilize the patient

in emergency situations.

Conformity Assessment: Class I medical device, self-certification by manufacturer; no requirement for NB

In accordance with Council Directive 93/42/EEC (MDD), Ferno-Washington, Inc. ("Ferno") declares the above named product(s) comply with the applicable provisions of the Medical Device Directive (MDD).

Ferno maintains an ISO 13485:2016 certification for its Quality Management System ensuring all medical devices are manufactured and distributed using consistent quality standards and post market surveillance and vigilance is maintained.

Compliance to additional standards/directives is noted as applicable:

This Declaration of Conformity is issued on this Loday of July, 2020 in Wilmington, Ohio, USA, under the sole responsibility of the manufacturer.

FERNO-WASHINGTON, INC.

By:

Dorothy Ramsey

Title:

VP, Global Legal & Regulatory

Signature: